

# ASO CONTROL

Lot xxxxxxxx



Cat. No.:	Pack name:	Packaging (Content):
BLT20004	ASO CON	1 x 1 ml

## Intended Use

Accuracy control for the determination of ASO in serum by turbidimetry and nephelometry.

## Composition

A dilution of human serum containing high levels of ASO with phosphate buffered saline.

Preservative: 0.095 g % sodium azide.

Ready to use.

## Storage and Stability

The expiry date of the product at 2-8 °C is listed on the label.

After first opening the container, the serum can be used for 6 weeks if stored tightly closed at 2-8 °C after use.

Do not freeze.

## Precautions and Warnings

1. For in Vitro Diagnostic use.
2. Each individual donation intended for use in manufacture of protein control serum was tested for hepatitis B surface antigen (HBsAg), anti-hepatitis C virus (anti-HCV) and anti-HIV1 and HIV 2 by FDA required test. Only donations with negative findings were used for its manufacture. Nevertheless every product obtained from human body fluids should be handled with appropriate care in accordance with recommended procedures for biohazardous materials since absence of infectious agents can never be proven.
3. Reagents containing sodium azide must be handled with due caution: Do not ingest or allow to contact skin or mucous membranes! Sodium azide can form explosive azides when contacting heavy metals such as copper or lead.

## Assigned Values

Titred against WHO standardisation: **xyz (.. - ...) IU/ml**

QUALITY SYSTEM CERTIFIED  
ISO 9001 ISO 13485



## SYMBOLS:


The following symbols are used in the labelling of ERBA kits:


 REF Catalogue No


 CE CE Mark - Device comply with the Directive 98/79/EC


 LOT Batch Code

 IVD In Vitro Diagnostics

 Expiry Date  
(Last day of the month)

 Consult Instruction for Use

 Manufactured by

 Storage temperature

 ASO CONTROL Product Name  CONT Content

*Date of Revision: 1.10.2013*

