

FERRITIN CALIBRATOR SET

Lot xxxxxxxx



Cat. No.:	Pack name:	Packaging (Content):
BLT20016	FERR CAL SET	5 x 1 ml

Intended Use

Preparation of reference curves for quantitative determination of Ferritin in human serum by turbidimetric immunoassay. The Calibrator Set has to be used with the Ferritin test kit.

Composition

Ferritin Calibrator Set are dilutions of human plasma, liquid stabilised and filtered through 0.2 μ . Preservative: Sodium azide (< 0.1 %).

Storage and Stability

Stability of unopened bottle 2-8°C: see bottle label.

After first opening the bottle the calibrator can be used for at least 6 weeks if stored tightly closed at 2-8 °C.

Precautions and Warnings

1. For *in Vitro* Diagnostic use only
2. Reagent is not classified as dangerous.
3. Each individual donation intended for use in manufacture of protein control serum was tested for hepatitis B surface antigen (HBsAg), anti-hepatitis C virus (anti-HCV) and anti-HIV1 and HIV 2 by CE certified tests. Only donations with negative findings were used for its manufacture. Nevertheless every product obtained from human body fluids should be handled with appropriate care in accordance with recommended procedures for biohazardous materials since absence of infectious agents can never be proven.
4. Reagents containing sodium azide must be handled with caution: Do not ingest or allow to contact skin or mucous membranes! Sodium azide can form explosive azides when contacting heavy metals such as copper or lead.

Assigned Values

	FER1 ng/ml	FER2 ng/ml	FER3 ng/ml	FER4 ng/ml	FER5 ng/ml
Ferritin	00.0	00.0	000.0	000.0	000.0

QUALITY SYSTEM CERTIFIED
ISO 9001 ISO 13485



SYMBOLS:

The following symbols are used in the labeling of ERBA kits:



Catalogue No



CE Mark - Device comply with the Directive 98/79/EC



Batch Code



In Vitro Diagnostics



Expiry Date



Consult Instruction for Use



Manufactured by



Storage temperature

FERRITIN CALIBRATOR SET

Product Name



Content

