

# FERRITIN CONTROL HIGH

Lot xxxxxxxx



| Cat. No.: | Pack name: | Packaging (Content): |
|-----------|------------|----------------------|
| BLT20017  | FERR CON H | 1 x 1 ml             |

## Intended Use

Accuracy control for the determination of FERRITIN in serum by turbidimetry and nephelometry.

## Composition

Ferritin control is a dilution of human plasma with phosphate buffered saline. Liquid stabilised.

Preservative: 0.1 % sodium azide.

Ready to use.

## Storage and Stability

The expiry date of the product at 2-8 °C is listed on the label.

After first opening the container, the serum can be used for 6 weeks if stored tightly closed at 2-8 °C after use.

Do not freeze.

## Precautions and Warnings

1. For *in Vitro* Diagnostic use.
2. Reagent is not classified as dangerous.
3. Each individual donation intended for use in manufacture of protein control serum was tested for hepatitis B surface antigen (HBsAg), anti-hepatitis C virus (anti-HCV) and anti-HIV1 and HIV 2 by FDA required test. Only donations with negative findings were used for its manufacture. Nevertheless every product obtained from human body fluids should be handled with appropriate care in accordance with recommended procedures for biohazardous materials since absence of infectious agents can never be proven.
4. Reagents containing sodium azide must be handled with due caution: Do not ingest or allow to contact skin or mucous membranes! Sodium azide can form explosive azides when contacting heavy metals such as copper or lead.

## Assigned Values

000 ng/ml (000 – 000)

QUALITY SYSTEM CERTIFIED  
ISO 9001 ISO 13485



**SYMBOLS:**

The following symbols are used in the labelling of ERBA kits:



Catalogue No



CE Mark - Device comply with the Directive 98/79/EC



Batch Code



In Vitro Diagnostics



Expiry Date



Consult Instruction for Use



Manufactured by



Storage temperature



Product name



Content

